

## TECHNICAL SPECIFICATIONS FOR ILLUMINATORS/NEGATOSCOPE/VIEWBOXES

- at least double bank
  - ◇ each bank with individual switches and dimmer
  - ◇ must be wall type
    - with diffused vertically oriented fluorescent lightning tube of uniform intensity, two 15 watts (linear type) bulbs per bank or panel
- with a diffusion panel (opal perspex screen)
- corrosion proof
- with shatter-proof diffusing panel
- with continuous film-retainer clip at top
- operational at 220 VAC, 60 Hz